

Certificate of Calibration

Certificate Number
240821 - 03870 - 9999999

REV: V1.00

Printed 08/21/2024

Instrument Identification

Company: 艾科尼斯股份有限公司

Address: 新北市汐止區大同路一段237號12樓-1

Model: 03870

Serial: 9999999

Customer Instrument ID : 9999999

Technician: Dennis Cheng

Order #:

Remarks: **As received, the instrument was found to be non-operational, and needed repair. The non-operational condition was due to corrupted firmware and output issues due to faulty parts. After service repair, calibration, and after test, the instrument is within manufacturer's tolerance limits.**

Received Date: **Aug, 21 / 2024**
Calibration Date: **Aug, 21 / 2024**
Issue Date: **Aug, 21 / 2024**
Recommended Due: **Aug, 20 / 2025**
Temperature: **(23+/- 5) DEG C**
Humidity: **(50+/-10) %RH**
Test Procedure #: **802-5001**

Calibration Location:
Ikonix


The instrument listed above has been calibrated using the standards traceable to the International System of Units (SI). Ikonix certifies that this instrument meets or exceeds manufacturer's tolerance limits as stated in the applicable test procedure (unless otherwise noted). The results reported herein relate only to the item calibrated. There was no sampling used in this calibration.

This certificate shall not be reproduced except in full, without the written approval of Ikonix.

Calibration Standards

Inst ID#	Model No	Model Description	Cal Date	Due Date	NIST Traceability
IKT001	HT-315	TEMPERATURE & HUMIDITY DATA LOGGER	2024/6/11	2025/6/10	OMEGA , PRT Temperature , PT100 , E-LAB-269 rotronic , Humidity and Temperature , HYGROPALM , 71908986
IKT002	VD15	HIGH VOLTAGE DIVIDER	2023/12/14	2024/12/13	Fluke , 8846A , UL(TAF1990) , ECT-Q23080802-01 PHENIX , KVM300 , NML(N0688) , E220097A PHENIX , KVM300 , TMCC(TAF3051) , C-20230715-01-04-A
IKT003	34465A	6.5 DIGIT MULTIMETER	2024/6/7	2025/6/6	Agilent , Digital Multimeter , 3458A , MY45045712 FLUKE , CALIBRATOR , 5500A , 9375028
IKT004	CTGB4-2	ATS EVOLUTION LOAD REV B	2024/1/24	2025/1/23	FLUKE , 52120A , 13052917-001 FLUKE , 8508A , 13040120-001 FLUKE , 5700A , 13031403-003 IET , HARS-X-8-0.01 , C3-11534562 IET , HRRS-B-6-1M , C1-1153446

Certified



Approved by: 